

Calibration Certificate

| | | | |
|------------------|--------------|-------------------|------------|
| Certificate No.: | TZ2025111808 | Certificate Date: | 2025/11/18 |
|------------------|--------------|-------------------|------------|

Reference Instrumentation

| Device | Manufacturer | Equipment No. | Accuracy | Calibration Date |
|---------------------------------------------|---------------------------------------|---------------|----------------|------------------|
| Programmable Temperature & Humidity Chamber | Donguan Kesonots Precision Instrument | KS-HW150 | ±0.05°C; ±3%RH | August,15th,2025 |
| High Accuracy RTD Thermometer | OMEGA | HH376 | ±0.15°C | August,15th,2025 |

Tzone Digital Technology Co., Ltd. certifies that the devices mentioned on this certificate have been thoroughly tested, validated and reached performance accuracy specifications over the specific ranges. These device were assembled, tested and calibrated in accordance with its specifications and prior to release for shipment. All reference instruments are calibrated by Shenzhen Zhongjidian Metrology and Testing Technology Co.,Ltd, a CNAS (China National Accreditation Service) accredited laboratory, and traceable to NIST according Mutual Recognition Agreement.

Product Information

| Product Name | Model | Serial No. | Production Date | Expiry Date | Quantity |
|----------------------------------------|---------|---------------------------|-----------------|-------------|----------|
| Disposable USB Temperature Data Logger | TempU08 | TZ2511000141-TZ2511000160 | 2025/11/18 | 2027/11/18 | 20 |

Calibration Results (Environment Conditions: Air Temperature 20.5°C, Relative Humidity 53.0%)

| Validation Points | Tolerance | Result |
|-------------------|-----------|--------|
| -20°C | ±0.5°C | Pass |
| 0°C | ±0.5°C | Pass |
| 30°C | ±0.5°C | Pass |

| | | | |
|-----------------------|------|-------------------------------------|---------------------------------------------------------------------------------------|
| Test Result Pass/Fail | Pass | Signature of Calibration Technician |  |
|-----------------------|------|-------------------------------------|---------------------------------------------------------------------------------------|

TZONE DIGITAL TECHNOLOGY COMPANY LIMITED

16D, Haiying building, South of Caitian Road, Futian District, Shenzhen City, China, 518033

Tel: 0086-755-82840647 Fax: 0086-755-82840648 Email: sales@tzonedigital.com Web: www.tzonedigital.com