

## Calibration Certificate

Certificate No.:	TZ2025111801	Certificate Date:	2025/11/18
------------------	--------------	-------------------	------------

### Reference Instrumentation

Device	Manufacturer	Equipment No.	Accuracy	Calibration Date
Programmable Temperature & Humidity Chamber	Donguan Kesonots Precision Instrument	KS-HW150	±0.05°C; ±3%RH	August,15th,2025
High Accuracy RTD Thermometer	OMEGA	HH376	±0.15°C	August,15th,2025
Low Temperature Thermostat	Jiangsu Tenlin Instrument	DC-8006	±0.1°C	February,25th,2025

Tzone Digital Technology Co., Ltd. certifies that the devices mentioned on this certificate have been thoroughly tested, validated and reached performance accuracy specifications over the specific ranges. These device were assembled, tested and calibrated in accordance with its specifications and prior to release for shipment. All reference instruments are calibrated by Shenzhen Zhongjidian Metrology and Testing Technology Co.,Ltd, a CNAS (China National Accreditation Service) accredited laboratories, and traceable to NIST according Mutual Recognition Agreement.

### Product Information

Product Name	Model	Serial No.	Production Date	Expiry Date	Quantity
Ultra Low Temperature Data Logger	TempU07 L100	TZ0725090331-TZ0725090368 TZ0725090370-TZ0725090378 TZ0725090380-TZ0725090382	2025/11/18	2026/11/18	50PCS

### Calibration Results (Environment Conditions: Air Temperature 20.5°C, Relative Humidity 53.0%)

Validation Points	Tolerance	Result
-80°C	±0.5°C	Pass
0°C	±0.5°C	Pass
60°C	±0.5°C	Pass

Test Result Pass/Fail	Pass	Signature of Calibration Technician
-----------------------	------	-------------------------------------



**TZONE DIGITAL TECHNOLOGY COMPANY LIMITED**

16D, Haiying building, South of Caijian Road, Futian District, Shenzhen City, China, 518033

Tel: 0086-755-82840647 Fax: 0086-755-82840648 Email: sales@tzonedit.com Web: www.tzonedit.com